
Midmark M11 - v1.0.0
Total Cycles: 1565

Name: RITTER 1
Sterilizer ID: V2178198
01/26/2022 10:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.4 F 0.0
2:00 175.7 F 2.9
4:00 192.6 F 6.9
6:00 203.5 F 11.9
8:00 220.2 F 5.7
10:00 229.7 F 10.5
12:00 231.7 F 7.2
14:00 243.6 F 12.9
16:00 242.1 F 11.2
18:00 254.9 F 17.7
20:00 266.1 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.6 F 28.9
2:00 271.8 F 28.1
3:00 272.3 F 28.8
4:00 272.5 F 28.7

Min 271.5 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.0 F 0.0
5:00 232.3 F 0.0
10:00 225.4 F 0.0
15:00 225.5 F 0.0
20:00 225.3 F 0.0
25:00 226.9 F 0.0
30:00 227.2 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 21:42
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:14
CYCLE COMPLETE

01/26/2022 11:54 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____