
Midmark M11 - v1.0.0
Total Cycles: 3280

Name: RITTER 1
Sterilizer ID: V2178198
05/24/2023 05:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.8 F 0.0
2:00 157.3 F 0.0
4:00 182.8 F 2.5
6:00 203.2 F 6.9
8:00 220.8 F 11.9
10:00 225.9 F 7.0
12:00 238.4 F 12.4
14:00 238.9 F 10.3
16:00 236.9 F 8.9
18:00 252.2 F 16.2
20:00 264.4 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.1
1:00 272.1 F 28.7
2:00 272.2 F 28.5
3:00 271.4 F 28.1
4:00 272.6 F 28.9

Min 271.4 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.7 F 0.0
5:00 229.0 F 0.0
10:00 231.4 F 0.0
15:00 229.3 F 0.0
20:00 227.4 F 0.0
25:00 228.8 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 21:46
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:00:09
CYCLE COMPLETE

05/24/2023 06:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____