
Midmark M11 - v1.0.0
Total Cycles: 3001

Name: RITTER 1
Sterilizer ID: V2178198
03/14/2023 09:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.5 F 0.0
2:00 174.7 F 2.2
4:00 181.4 F 5.7
6:00 204.8 F 10.2
8:00 216.9 F 5.0
10:00 230.5 F 10.1
12:00 233.3 F 8.1
14:00 243.4 F 13.0
16:00 242.5 F 11.7
18:00 255.5 F 18.4
20:00 266.8 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.5 F 29.1
2:00 271.4 F 28.4
3:00 272.7 F 29.3
4:00 271.8 F 28.6

Min 271.3 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.4 F 0.0
5:00 210.6 F 0.0
10:00 217.5 F 0.0
15:00 223.8 F 0.0
20:00 231.9 F 0.0
25:00 226.9 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 21:29
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:48
CYCLE COMPLETE

03/14/2023 10:19 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____