

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000195  
8:56 18/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 0:55  
PRESSURIZING 0:58  
STERILIZING 1:58  
136.4 C 232kPa 1:58  
136.2 C 231kPa 2:28  
136.4 C 233kPa 2:58  
136.3 C 232kPa 3:28  
136.4 C 232kPa 3:58  
136.4 C 230kPa 4:28  
136.4 C 231kPa 4:58  
136.4 C 233kPa 5:28  
136.4 C 233kPa 5:28  
VENTING 5:28  
AIR DRYING 6:16  
CYCLE COMPLETE 6:45  
-----

Digital Signature #  
E96A56FF5C88BB6B  
-----