
Midmark M11 - v1.0.5
Total Cycles: 2594

Name: RITTER2
Sterilizer ID: V2313228
06/30/2023 09:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.4 F 0.0
2:00 198.1 F 3.3
4:00 204.5 F 8.2
6:00 227.1 F 11.2
8:00 229.7 F 8.6
10:00 240.2 F 11.5
12:00 241.6 F 12.1
14:00 241.0 F 11.1
16:00 254.0 F 17.7
18:00 265.0 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.7 F 29.5
2:00 271.8 F 28.7
3:00 272.2 F 29.5
4:00 272.5 F 29.3

Min 271.5 F 28.7
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.9 F 0.0
5:00 204.1 F 0.0
10:00 214.1 F 0.0
15:00 218.9 F 0.0
20:00 217.7 F 0.0
25:00 213.8 F 0.0
30:00 211.2 F 0.0
Drying Complete

FILLING: 2:07
HEATING: 19:55
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 00:59:24
CYCLE COMPLETE

06/30/2023 10:56 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____