
Midmark M11 - v1.0.5
Total Cycles: 78

Name: RITTER2
Sterilizer ID: V2313228
06/08/2021 01:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.4 F 0.0
2:00 134.7 F 1.7
4:00 149.2 F 4.6
6:00 175.4 F 8.3
8:00 196.1 F 12.5
10:00 208.0 F 5.5
12:00 221.9 F 10.1
14:00 229.7 F 8.7
16:00 239.3 F 12.6
18:00 240.3 F 11.6
20:00 252.6 F 17.7
22:00 263.9 F 24.1
24:00 272.5 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.8
1:00 272.6 F 30.1
2:00 272.1 F 30.2
3:00 272.3 F 30.0
4:00 272.1 F 30.2

Min 271.7 F 29.6
Max 273.2 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.1 F 0.0
5:00 221.7 F 0.0
10:00 233.4 F 0.0
15:00 227.1 F 0.0
20:00 222.6 F 0.0
25:00 234.3 F 0.0
30:00 220.4 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 24:12
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:02:44
CYCLE COMPLETE

06/08/2021 02:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____