

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2980

Name: RITTER2  
Sterilizer ID: V2313228  
10/05/2023 01:38 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 164.4 F 0.0  
2:00 165.7 F 2.8  
4:00 184.2 F 6.7  
6:00 202.7 F 11.2  
8:00 219.4 F 5.5  
10:00 228.9 F 10.5  
12:00 232.2 F 8.0  
14:00 244.7 F 13.5  
16:00 243.7 F 12.5  
18:00 256.0 F 18.9  
20:00 266.7 F 25.6  
22:00 272.8 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 272.4 F 29.3  
1:00 272.4 F 29.6  
2:00 272.2 F 29.8  
3:00 271.6 F 29.5  
4:00 272.3 F 29.4

Min 271.5 F 29.2  
Max 272.6 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.5 F 0.0  
5:00 227.0 F 0.0  
10:00 215.1 F 0.0  
15:00 203.5 F 0.0  
20:00 178.3 F 0.0  
25:00 174.9 F 0.0  
30:00 173.5 F 0.0  
Drying Complete

FILLING: 1:39  
HEATING: 22:16  
STERILIZING: 4:00  
VENTING: 3:34  
DRYING: 30:00  
TOTAL CYCLE: 01:01:28  
CYCLE COMPLETE

10/05/2023 02:39 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_