
Midmark M11 - v1.0.5
Total Cycles: 1625

Name: RITTER2
Sterilizer ID: V2313228
10/19/2022 01:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.1 F 0.0
2:00 181.0 F 3.7
4:00 192.8 F 8.2
6:00 214.5 F 12.5
8:00 228.5 F 8.9
10:00 236.1 F 10.3
12:00 240.3 F 12.2
14:00 240.5 F 11.4
16:00 254.4 F 18.4
18:00 266.5 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.4 F 29.7
2:00 271.6 F 29.6
3:00 272.2 F 29.7
4:00 271.3 F 29.3

Min 271.3 F 29.1
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.9 F 0.0
5:00 215.3 F 0.0
10:00 225.4 F 0.0
15:00 247.8 F 0.0
20:00 235.3 F 0.0
25:00 236.7 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 2:04
HEATING: 19:29
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:49
CYCLE COMPLETE

10/19/2022 02:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____