
Midmark M11 - v1.0.0
Total Cycles: 1672

Name: RITTER 1
Sterilizer ID: V2178198
02/24/2022 03:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.2 F 0.0
2:00 177.7 F 2.5
4:00 190.7 F 6.7
6:00 216.9 F 12.1
8:00 227.8 F 7.0
10:00 239.2 F 12.3
12:00 237.9 F 9.6
14:00 237.9 F 8.7
16:00 249.4 F 14.6
18:00 261.4 F 21.4
20:00 272.5 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.0
1:00 272.2 F 28.6
2:00 271.2 F 27.9
3:00 272.6 F 28.8
4:00 272.1 F 28.1

Min 271.1 F 27.8
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.8 F 0.0
5:00 229.4 F 0.0
10:00 212.3 F 0.0
15:00 209.8 F 0.0
20:00 215.2 F 0.0
25:00 206.7 F 0.0
30:00 212.1 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 20:27
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:59:18
CYCLE COMPLETE

02/24/2022 05:03 PM
ACCEPTED: X
REJECTED:
Operator: U02