
Midmark M11 - v1.0.5
Total Cycles: 1684

Name: RITTER2
Sterilizer ID: V2313228
11/04/2022 09:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.8 F 0.0
2:00 126.1 F 1.2
4:00 141.1 F 4.0
6:00 171.1 F 8.2
8:00 193.9 F 13.1
10:00 222.6 F 6.8
12:00 238.3 F 13.0
14:00 239.2 F 10.9
16:00 238.1 F 9.7
18:00 253.1 F 17.1
20:00 265.8 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.0 F 28.6
2:00 272.7 F 29.1
3:00 272.7 F 29.5
4:00 271.7 F 28.5

Min 271.6 F 28.4
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 222.1 F 0.0
Dry Cycle Aborted

FILLING: 1:12
HEATING: 21:32
STERILIZING: 4:00
VENTING: 2:42
DRYING: 2:21
TOTAL CYCLE: 00:31:48

11/04/2022 09:58 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____