
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001481
7:30 10/07/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:07
PRESSURIZING 6:58
STERILIZING 8:47
136.0 C 220kPa 8:47
136.1 C 218kPa 9:17
136.3 C 221kPa 9:47
136.0 C 218kPa 10:17
136.4 C 222kPa 10:47
136.1 C 220kPa 11:17
136.0 C 218kPa 11:47
136.4 C 222kPa 12:17
VENTING 12:17
AIR DRYING 13:16
CYCLE COMPLETE 43:17

Digital Signature #
92484C44E45C3DDC
