
Midmark M11 - v1.0.0
Total Cycles: 5353

Name: RITTER 1
Sterilizer ID: V2178198
02/21/2025 09:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.1 F 0.0
2:00 90.1 F 0.3
4:00 103.9 F 1.8
6:00 122.4 F 4.3
8:00 154.3 F 7.2
10:00 187.7 F 10.5
12:00 199.2 F 2.6
14:00 214.3 F 5.3
16:00 223.1 F 9.0
18:00 233.5 F 13.1
20:00 234.0 F 8.6
22:00 242.3 F 11.7
24:00 243.8 F 12.0
26:00 256.4 F 18.6
28:00 267.7 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.4 F 28.5
2:00 272.3 F 28.7
3:00 271.8 F 28.2
4:00 272.7 F 28.9

Min 271.5 F 28.0
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.5 F 0.0
5:00 191.5 F 0.0
10:00 200.6 F 0.0
15:00 197.7 F 0.0
20:00 196.1 F 0.0
25:00 202.3 F 0.0
30:00 208.0 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 29:20
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:07:32
CYCLE COMPLETE

02/21/2025 10:14 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____