

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001930  
10:14 08/08/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:27  
PRESSURIZING 1:34  
STERILIZING 2:30  
136.0 C 220kPa 2:30  
136.2 C 221kPa 3:00  
136.4 C 226kPa 3:30  
136.4 C 225kPa 4:00  
136.0 C 222kPa 4:30  
136.3 C 224kPa 5:00  
136.4 C 226kPa 5:30  
136.3 C 224kPa 6:00  
136.3 C 224kPa 6:00  
VENTING 6:00  
AIR DRYING 6:52  
CYCLE COMPLETE 30:50  
-----

Digital Signature #  
A1A89BF9084F4D0D  
-----