
Midmark M11 - v1.0.5
Total Cycles: 1053

Name: RITTER2
Sterilizer ID: V2313228
05/02/2022 09:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.2 F 0.0
2:00 179.3 F 2.9
4:00 201.6 F 7.8
6:00 216.4 F 12.8
8:00 226.9 F 8.7
10:00 237.1 F 10.8
12:00 242.5 F 12.8
14:00 242.7 F 12.2
16:00 255.6 F 18.9
18:00 266.8 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.7 F 29.8
2:00 271.5 F 29.1
3:00 272.5 F 29.9
4:00 271.8 F 29.3

Min 271.3 F 29.1
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.8 F 0.0
5:00 211.3 F 0.0
10:00 218.4 F 0.0
15:00 215.8 F 0.0
20:00 209.5 F 0.0
25:00 207.1 F 0.0
30:00 209.9 F 0.0
Drying Complete

FILLING: 3:13
HEATING: 19:31
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:00:01
CYCLE COMPLETE

05/02/2022 10:38 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____