
Midmark M11 - v1.0.0
Total Cycles: 4626

Name: RITTER 1
Sterilizer ID: V2178198
06/11/2024 01:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 86.8 F 0.0
2:00 92.6 F 0.6
4:00 119.6 F 2.5
6:00 140.2 F 4.9
8:00 166.6 F 7.9
10:00 185.8 F 11.6
12:00 198.4 F 2.5
14:00 211.1 F 5.7
16:00 221.8 F 9.1
18:00 231.4 F 12.9
20:00 232.0 F 8.5
22:00 243.2 F 13.4
24:00 241.7 F 11.3
26:00 252.8 F 16.9
28:00 263.0 F 22.8
30:00 272.6 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.8 F 29.1
2:00 271.8 F 28.5
3:00 271.8 F 28.8
4:00 272.7 F 29.1

Min 271.4 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 205.9 F 0.0
10:00 194.5 F 0.0
15:00 199.3 F 0.0
20:00 215.1 F 0.0
25:00 216.0 F 0.0
30:00 217.4 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 30:26
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:09:03
CYCLE COMPLETE

06/11/2024 02:34 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____