
Midmark M11 - v1.0.0
Total Cycles: 4843

Name: RITTER 1
Sterilizer ID: V2178198
08/22/2024 12:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 111.7 F 0.0
2:00 112.4 F 0.1
4:00 126.2 F 0.8
6:00 159.4 F 1.3
8:00 188.1 F 3.8
10:00 197.0 F 7.3
12:00 210.6 F 10.8
14:00 222.5 F 6.1
16:00 226.8 F 8.7
18:00 237.9 F 13.2
20:00 238.1 F 10.0
22:00 238.7 F 9.0
24:00 247.3 F 13.7
26:00 258.0 F 19.6
28:00 268.0 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.1
1:00 272.3 F 28.9
2:00 272.0 F 28.5
3:00 271.2 F 28.1
4:00 272.1 F 28.9

Min 271.1 F 28.0
Max 272.4 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.9 F 0.0
5:00 192.6 F 0.0
10:00 184.0 F 0.0
15:00 185.1 F 0.0
20:00 200.0 F 0.0
25:00 209.9 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 29:28
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:07:57
CYCLE COMPLETE

08/22/2024 01:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____