
Midmark M11 - v1.0.0
Total Cycles: 937

Name: RITTER 1
Sterilizer ID: V2178198
06/24/2021 12:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.6 F 0.0
2:00 166.2 F 2.9
4:00 186.5 F 7.9
6:00 212.9 F 10.7
8:00 219.4 F 7.9
10:00 233.8 F 13.5
12:00 240.3 F 11.0
14:00 238.1 F 9.2
16:00 251.8 F 16.1
18:00 263.6 F 23.0
20:00 271.5 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.9 F 29.5
2:00 271.9 F 28.8
3:00 272.6 F 29.7
4:00 272.8 F 29.3

Min 271.5 F 28.8
Max 273.1 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.0 F 0.0
5:00 232.0 F 0.0
10:00 235.8 F 0.0
15:00 230.8 F 0.0
20:00 229.3 F 0.0
25:00 227.9 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 20:54
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 00:59:18
CYCLE COMPLETE

06/24/2021 01:25 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____