
Midmark M11 - v1.0.0
Total Cycles: 4697

Name: RITTER 1
Sterilizer ID: V2178198
07/03/2024 02:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 94.5 F 0.0
2:00 95.6 F 0.7
4:00 106.8 F 2.8
6:00 135.8 F 5.5
8:00 162.7 F 8.6
10:00 183.5 F 12.2
12:00 201.8 F 4.0
14:00 214.4 F 7.6
16:00 227.2 F 11.7
18:00 227.6 F 7.5
20:00 240.5 F 12.7
22:00 239.3 F 10.6
24:00 251.7 F 16.4
26:00 261.7 F 22.5
28:00 271.8 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.6 F 29.4
2:00 271.2 F 28.7
3:00 272.5 F 29.5
4:00 271.7 F 28.9

Min 271.2 F 28.7
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.5 F 0.0
5:00 227.5 F 0.0
10:00 235.2 F 0.0
15:00 228.3 F 0.0
20:00 229.7 F 0.0
25:00 229.7 F 0.0
30:00 231.7 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 28:34
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:07:09
CYCLE COMPLETE

07/03/2024 03:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____