

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001070  
19:33 10/05/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:43  
PRESSURIZING 4:44  
STERILIZING 5:57  
136.3 C 220kPa 5:57  
136.4 C 225kPa 6:27  
136.3 C 224kPa 6:57  
136.4 C 225kPa 7:27  
136.1 C 223kPa 7:57  
136.3 C 224kPa 8:27  
136.4 C 225kPa 8:57  
136.3 C 223kPa 9:27  
136.3 C 223kPa 9:27  
VENTING 9:27  
AIR DRYING 10:26  
CYCLE COMPLETE 25:26  
-----

Digital Signature #  
A178A726B72B5049  
-----