

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1308

Name: RITTER 1  
Sterilizer ID: V2178198  
10/27/2021 11:16 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 183.2 F 0.0  
2:00 179.7 F 2.6  
4:00 195.9 F 6.1  
6:00 218.3 F 10.7  
8:00 225.2 F 5.3  
10:00 234.1 F 9.9  
12:00 239.2 F 9.1  
14:00 243.4 F 12.6  
16:00 241.9 F 11.3  
18:00 254.6 F 17.7  
20:00 265.6 F 24.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.7 F 29.1  
2:00 271.6 F 28.3  
3:00 272.4 F 29.2  
4:00 272.1 F 28.7

Min 271.3 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.9 F 0.0  
5:00 229.1 F 0.0  
10:00 229.5 F 0.0  
15:00 227.2 F 0.0  
20:00 226.8 F 0.0  
25:00 229.5 F 0.0  
30:00 228.7 F 0.0  
Drying Complete

FILLING: 1:50  
HEATING: 21:47  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 01:01:04  
CYCLE COMPLETE

10/27/2021 12:33 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01