

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001562  
13:30 06/11/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:17  
PRESSURIZING 3:23  
STERILIZING 4:33  
136.3 C 223kPa 4:33  
136.3 C 224kPa 5:03  
136.3 C 224kPa 5:33  
136.1 C 222kPa 6:03  
136.4 C 226kPa 6:33  
136.4 C 226kPa 7:03  
136.3 C 225kPa 7:33  
136.4 C 226kPa 8:03  
136.4 C 226kPa 8:03  
VENTING 8:03  
AIR DRYING 8:58  
CYCLE COMPLETE 9:13  
-----

Digital Signature #  
CFB0E2CF0F74F753  
-----