

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1918

Name: RITTER2  
Sterilizer ID: V2313228  
01/11/2023 10:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 192.5 F 0.0  
2:00 192.1 F 3.1  
4:00 196.5 F 8.1  
6:00 211.0 F 10.6  
8:00 224.0 F 8.5  
10:00 235.9 F 11.1  
12:00 241.4 F 12.2  
14:00 241.4 F 11.3  
16:00 255.1 F 18.4  
18:00 267.2 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.5 F 29.5  
2:00 271.4 F 28.9  
3:00 272.4 F 29.6  
4:00 271.8 F 29.0  
  
Min 271.3 F 28.8  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.3 F 0.0  
5:00 226.5 F 0.0  
10:00 228.7 F 0.0  
15:00 230.8 F 0.0  
20:00 229.3 F 0.0  
25:00 228.4 F 0.0  
30:00 229.9 F 0.0  
Drying Complete

FILLING: 1:33  
HEATING: 19:24  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 00:58:02  
CYCLE COMPLETE

01/11/2023 11:55 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_