
Midmark M11 - v1.0.5
Total Cycles: 2476

Name: RITTER2
Sterilizer ID: V2313228
06/07/2023 07:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.9 F 0.0
2:00 78.2 F 0.0
4:00 90.6 F 0.0
6:00 111.7 F 0.0
8:00 151.0 F 1.7
10:00 174.2 F 5.1
12:00 203.1 F 8.8
14:00 219.2 F 12.7
16:00 223.1 F 6.3
18:00 232.7 F 10.5
20:00 232.5 F 8.3
22:00 242.1 F 12.7
24:00 241.1 F 11.2
26:00 253.2 F 17.4
28:00 263.9 F 23.8
30:00 271.3 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.9
1:00 272.3 F 29.5
2:00 271.1 F 29.0
3:00 272.2 F 29.5
4:00 271.2 F 28.8

Min 271.0 F 28.7
Max 272.3 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.3 F 0.0
5:00 203.9 F 0.0
10:00 210.8 F 0.0
15:00 218.7 F 0.0
20:00 213.8 F 0.0
25:00 219.0 F 0.0
30:00 218.4 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:05
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:08:03
CYCLE COMPLETE

06/07/2023 08:35 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____