

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2362

Name: RITTER 1  
Sterilizer ID: V2178198  
09/22/2022 02:08 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 191.5 F 0.0  
2:00 190.2 F 4.2  
4:00 198.2 F 9.0  
6:00 210.7 F 8.9  
8:00 223.9 F 8.9  
10:00 234.7 F 9.9  
12:00 240.6 F 11.6  
14:00 238.7 F 9.9  
16:00 252.6 F 16.6  
18:00 264.0 F 23.4  
20:00 271.2 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.2  
1:00 272.7 F 29.4  
2:00 272.3 F 29.7  
3:00 272.2 F 29.1  
4:00 272.7 F 29.7  
  
Min 271.8 F 29.0  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.0 F 0.0  
5:00 219.1 F 0.0  
10:00 230.9 F 0.0  
15:00 233.8 F 0.0  
20:00 226.0 F 0.0  
25:00 226.0 F 0.0  
30:00 225.6 F 0.0  
Drying Complete

FILLING: 1:37  
HEATING: 20:54  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:59:45  
CYCLE COMPLETE

09/22/2022 03:07 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_