

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001464  
11:17 20/09/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:00  
PRESSURIZING 2:05  
STERILIZING 3:03  
136.4 C 225kPa 3:03  
136.4 C 224kPa 3:33  
136.4 C 226kPa 4:03  
136.4 C 226kPa 4:33  
136.2 C 221kPa 5:03  
136.4 C 225kPa 5:33  
136.2 C 226kPa 6:03  
136.4 C 224kPa 6:33  
136.4 C 224kPa 6:33  
VENTING 6:33  
AIR DRYING 7:24  
CYCLE COMPLETE 7:28  
-----

Digital Signature #  
C6A020AEB05B3901  
-----