
Midmark M11 - v1.0.5
Total Cycles: 3711

Name: RITTER2
Sterilizer ID: V2313228
06/18/2024 11:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.3 F 0.0
2:00 167.3 F 3.3
4:00 175.6 F 7.1
6:00 198.5 F 12.0
8:00 215.5 F 6.3
10:00 231.2 F 11.7
12:00 233.0 F 8.8
14:00 239.3 F 10.2
16:00 246.5 F 13.9
18:00 258.9 F 20.7
20:00 270.0 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.6 F 29.6
2:00 271.5 F 29.1
3:00 272.5 F 29.7
4:00 271.6 F 28.9

Min 271.4 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.8 F 0.0
5:00 216.0 F 0.0
10:00 218.0 F 0.0
15:00 213.9 F 0.0
20:00 210.6 F 0.0
25:00 212.1 F 0.0
30:00 209.7 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 20:56
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:59:21
CYCLE COMPLETE

06/18/2024 12:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____