
Midmark M11 - v1.0.5
Total Cycles: 3563

Name: RITTER2
Sterilizer ID: V2313228
04/26/2024 09:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.2 F 0.0
2:00 148.2 F 1.3
4:00 155.3 F 3.9
6:00 178.4 F 7.6
8:00 197.3 F 11.8
10:00 222.7 F 5.3
12:00 233.3 F 10.0
14:00 235.9 F 9.0
16:00 240.5 F 12.0
18:00 239.9 F 10.3
20:00 251.8 F 16.3
22:00 262.0 F 22.4
24:00 272.1 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.3 F 29.2
2:00 272.0 F 29.7
3:00 271.2 F 29.1
4:00 272.5 F 29.4

Min 271.1 F 29.0
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.1 F 0.0
5:00 239.9 F 0.0
10:00 209.5 F 0.0
15:00 178.7 F 0.0
20:00 175.8 F 0.0
25:00 173.6 F 0.0
30:00 178.9 F 0.0
Drying Complete

FILLING: 1:29
HEATING: 25:22
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:04:20
CYCLE COMPLETE

04/26/2024 11:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____