

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3743

Name: RITTER 1  
Sterilizer ID: V2178198  
09/08/2023 01:27 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 174.6 F 0.0  
2:00 173.5 F 1.0  
4:00 191.9 F 4.0  
6:00 207.4 F 8.2  
8:00 221.7 F 12.9  
10:00 233.4 F 8.4  
12:00 244.4 F 13.5  
14:00 241.5 F 11.3  
16:00 238.9 F 9.8  
18:00 252.1 F 16.3  
20:00 263.2 F 22.7  
22:00 272.3 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.4 F 28.5  
2:00 272.0 F 28.9  
3:00 272.1 F 28.7  
4:00 272.6 F 29.2

Min 270.9 F 28.1  
Max 273.0 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.5 F 0.0  
5:00 197.0 F 0.0  
10:00 198.8 F 0.0  
15:00 195.4 F 0.0  
20:00 191.0 F 0.0  
25:00 186.4 F 0.0  
30:00 182.3 F 0.0  
Drying Complete

FILLING: 2:05  
HEATING: 22:13  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 01:01:53  
CYCLE COMPLETE

09/08/2023 02:28 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_