
Midmark M11 - v1.0.5
Total Cycles: 4487

Name: RITTER2
Sterilizer ID: V2313228
04/04/2025 10:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.1 F 0.0
2:00 136.8 F 1.3
4:00 139.8 F 3.9
6:00 156.3 F 7.5
8:00 181.5 F 11.8
10:00 211.0 F 4.7
12:00 224.8 F 8.8
14:00 230.5 F 13.1
16:00 233.4 F 9.9
18:00 236.3 F 9.2
20:00 245.7 F 13.9
22:00 256.9 F 20.0
24:00 267.4 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.6
1:00 272.4 F 30.0
2:00 271.3 F 29.9
3:00 272.3 F 29.9
4:00 271.3 F 30.0

Min 271.1 F 29.5
Max 272.6 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.4 F 0.0
5:00 200.7 F 0.0
10:00 193.3 F 0.0
15:00 190.6 F 0.0
20:00 193.0 F 0.0
25:00 193.3 F 0.0
30:00 196.6 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 25:32
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:04:17
CYCLE COMPLETE

04/04/2025 11:43 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____