
Midmark M11 - v1.0.0
Total Cycles: 2175

Name: RITTER 1
Sterilizer ID: V2178198
07/26/2022 08:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.2 F 0.0
2:00 134.9 F 0.0
4:00 146.8 F 2.1
6:00 175.4 F 6.0
8:00 207.1 F 11.1
10:00 223.5 F 4.9
12:00 229.8 F 9.8
14:00 237.0 F 8.8
16:00 241.0 F 11.5
18:00 237.9 F 9.3
20:00 251.6 F 16.0
22:00 263.1 F 22.6
24:00 272.5 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.5 F 28.9
2:00 271.5 F 28.5
3:00 272.6 F 29.0
4:00 271.5 F 28.3

Min 271.4 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.6 F 0.0
5:00 211.8 F 0.0
10:00 184.8 F 0.0
15:00 188.0 F 0.0
20:00 216.3 F 0.0
25:00 219.3 F 0.0
30:00 239.5 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 24:60
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:03:15
CYCLE COMPLETE

07/26/2022 09:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____