
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.7uS / 0.4ppm
CYCLE NUMBER 003660
9:47 12/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:34
STERILIZING 3:31
136.5 C 221kPa 3:31
136.3 C 223kPa 4:01
136.4 C 224kPa 4:31
136.4 C 222kPa 5:01
136.1 C 221kPa 5:31
136.4 C 222kPa 6:01
136.3 C 224kPa 6:31
136.4 C 222kPa 7:01
136.4 C 222kPa 7:01
VENTING 7:01
AIR DRYING 7:53
CYCLE COMPLETE 7:59

Digital Signature #
5A45B063A9F9D3FE
