
Midmark M11 - v1.0.5
Total Cycles: 3228

Name: RITTER2
Sterilizer ID: V2313228
12/18/2023 12:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.7 F 0.0
2:00 131.4 F 1.0
4:00 152.1 F 3.6
6:00 178.8 F 7.6
8:00 196.5 F 12.3
10:00 221.7 F 6.4
12:00 232.5 F 11.1
14:00 229.6 F 7.5
16:00 243.4 F 13.3
18:00 241.4 F 11.9
20:00 254.5 F 18.5
22:00 265.9 F 25.5
24:00 273.0 F 30.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.7
1:00 272.6 F 29.8
2:00 272.7 F 30.2
3:00 271.7 F 30.0
4:00 272.4 F 29.8

Min 271.5 F 29.5
Max 272.9 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.7 F 0.0
5:00 227.0 F 0.0
10:00 192.8 F 0.0
15:00 187.0 F 0.0
20:00 184.7 F 0.0
25:00 183.4 F 0.0
30:00 181.2 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 24:19
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:02:30
CYCLE COMPLETE

12/18/2023 01:45 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____