

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001320  
15:14 26/10/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:31  
PRESSURIZING 3:33  
STERILIZING 4:56  
136.0 C 221kPa 4:56  
136.4 C 224kPa 5:26  
136.4 C 224kPa 5:56  
136.2 C 223kPa 6:26  
136.4 C 221kPa 6:56  
136.1 C 223kPa 7:26  
136.4 C 225kPa 7:56  
136.4 C 225kPa 8:26  
136.4 C 225kPa 8:26  
VENTING 8:27  
AIR DRYING 9:24  
CYCLE COMPLETE 16:06  
-----

Digital Signature #  
26B29D91F7AD0C7E  
-----