

-----  
Mississauga, ON  
Canada  
-----

STATIM 5000 S501R751  
SN 511020J00004  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000824  
12:09 30/03/2022  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 5:16  
PRESSURIZING 7:08  
STERILIZING 8:43  
136.0 C 220kPa 8:43  
136.1 C 222kPa 9:13  
136.4 C 223kPa 9:43  
136.4 C 222kPa 10:13  
136.4 C 223kPa 10:43  
136.2 C 221kPa 11:13  
136.4 C 223kPa 11:43  
136.4 C 223kPa 12:13  
136.4 C 223kPa 12:13  
VENTING 12:13  
AIR DRYING 13:17  
CYCLE COMPLETE 43:17  
-----

Digital Signature #  
BE246E249B6C7621  
-----