

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4004

Name: RITTER2  
Sterilizer ID: V2313228  
10/07/2024 08:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 67.9 F 0.0  
2:00 71.2 F 0.0  
4:00 95.8 F 0.0  
6:00 109.0 F 0.1  
8:00 145.0 F 0.2  
10:00 173.7 F 0.7  
12:00 194.0 F 3.8  
14:00 206.2 F 7.1  
16:00 218.6 F 10.8  
18:00 225.5 F 7.2  
20:00 232.5 F 9.9  
22:00 236.6 F 9.7  
24:00 242.9 F 12.5  
26:00 241.2 F 11.3  
28:00 253.4 F 17.4  
30:00 263.8 F 23.6  
32:00 272.0 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 272.5 F 29.3  
2:00 271.8 F 29.1  
3:00 272.3 F 29.2  
4:00 271.8 F 29.2  
  
Min 271.5 F 28.6  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.0 F 0.0  
5:00 193.6 F 0.0  
10:00 201.5 F 0.0  
15:00 200.7 F 0.0  
20:00 201.5 F 0.0  
25:00 198.4 F 0.0  
30:00 198.0 F 0.0  
  
Drying Complete

FILLING: 0:50  
HEATING: 32:08  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:10:24  
CYCLE COMPLETE

10/07/2024 09:57 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_