

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5280

Name: RITTER 1  
Sterilizer ID: V2178198  
01/27/2025 01:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 120.5 F 0.0  
2:00 127.4 F 0.1  
4:00 137.2 F 2.2  
6:00 162.2 F 5.3  
8:00 184.4 F 8.9  
10:00 200.6 F 12.7  
12:00 213.4 F 4.5  
14:00 224.4 F 8.4  
16:00 235.0 F 12.6  
18:00 234.0 F 9.3  
20:00 240.8 F 10.6  
22:00 244.8 F 13.1  
24:00 255.5 F 18.9  
26:00 265.7 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.9  
1:00 272.2 F 29.5  
2:00 272.7 F 29.7  
3:00 272.6 F 29.5  
4:00 272.4 F 29.2

Min 271.5 F 28.8  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.0 F 0.0  
5:00 211.9 F 0.0  
10:00 200.0 F 0.0  
15:00 201.8 F 0.0  
20:00 210.7 F 0.0  
25:00 205.7 F 0.0  
30:00 206.8 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 27:57  
STERILIZING: 4:00  
VENTING: 3:60  
DRYING: 30:00  
TOTAL CYCLE: 01:06:45  
CYCLE COMPLETE

01/27/2025 02:21 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_