
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002381
7:32 07/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:35
PRESSURIZING 2:40
STERILIZING 3:28
136.0 C 220kPa 3:28
135.9 C 220kPa 3:58
136.2 C 223kPa 4:28
135.9 C 218kPa 4:58
136.4 C 220kPa 5:28
136.2 C 221kPa 5:58
136.0 C 218kPa 6:28
136.3 C 222kPa 6:58
136.3 C 222kPa 6:58
VENTING 6:58
AIR DRYING 7:47
CYCLE COMPLETE 9:15

Digital Signature #
0BF040A0FBCD229D
