
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000236
9:05 20/07/2020

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:02
PRESSURIZING 3:07
STERILIZING 4:05
136.5 C 224kPa 4:05
136.2 C 222kPa 4:35
136.3 C 223kPa 5:05
136.4 C 224kPa 5:35
136.3 C 223kPa 6:05
136.3 C 222kPa 6:35
136.0 C 222kPa 7:05
136.4 C 223kPa 7:35
136.4 C 223kPa 7:35
VENTING 7:35
AIR DRYING 8:25
CYCLE COMPLETE 8:28

Digital Signature #
45589EDB7F163F32
