
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001219
12:43 19/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:44
PRESSURIZING 3:44
STERILIZING 4:44
136.4 C 223kPa 4:44
136.3 C 224kPa 5:14
136.4 C 224kPa 5:44
136.5 C 223kPa 6:14
136.4 C 224kPa 6:44
136.2 C 223kPa 7:14
136.3 C 224kPa 7:44
136.2 C 221kPa 8:14
136.2 C 221kPa 8:14
VENTING 8:14
AIR DRYING 9:11
CYCLE COMPLETE 9:15

Digital Signature #
51511226BD4D73DB
