
Midmark M11 - v1.0.0
Total Cycles: 2128

Name: RITTER 1
Sterilizer ID: V2178198
07/14/2022 11:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.3 F 0.0
2:00 157.9 F 0.1
4:00 182.0 F 4.3
6:00 200.3 F 9.5
8:00 226.6 F 6.8
10:00 231.6 F 8.6
12:00 243.9 F 13.0
14:00 240.4 F 10.7
16:00 235.8 F 8.4
18:00 250.2 F 15.1
20:00 261.5 F 21.6
22:00 272.1 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.6
1:00 271.9 F 28.6
2:00 271.8 F 28.6
3:00 271.8 F 28.5
4:00 271.8 F 28.5

Min 271.8 F 28.5
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.5 F 0.0
5:00 233.0 F 0.0
10:00 222.3 F 0.0
15:00 228.5 F 0.0
20:00 205.5 F 0.0
25:00 214.5 F 0.0
30:00 161.4 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 23:05
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:01:10
CYCLE COMPLETE

07/14/2022 12:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____