
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002035
13:18 14/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:45
PRESSURIZING 2:52
STERILIZING 3:55
136.4 C 223kPa 3:55
136.2 C 223kPa 4:25
136.3 C 221kPa 4:55
136.4 C 225kPa 5:25
135.8 C 218kPa 5:55
136.3 C 222kPa 6:25
136.2 C 223kPa 6:55
136.5 C 224kPa 7:25
136.5 C 224kPa 7:25
VENTING 7:25
AIR DRYING 8:17
CYCLE COMPLETE 8:20

Digital Signature #
38D7929ACF243C60
