

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4177

Name: RITTER 1  
Sterilizer ID: V2178198  
01/03/2024 03:45 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 161.9 F 0.0  
2:00 158.9 F 2.3  
4:00 166.9 F 5.9  
6:00 201.3 F 9.4  
8:00 221.4 F 7.5  
10:00 225.3 F 7.6  
12:00 237.1 F 12.2  
14:00 234.9 F 8.9  
16:00 241.2 F 10.7  
18:00 245.5 F 13.0  
20:00 257.1 F 19.2  
22:00 267.6 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.4  
1:00 272.2 F 28.9  
2:00 271.4 F 28.8  
3:00 272.2 F 28.8  
4:00 271.4 F 28.8

Min 271.0 F 28.3  
Max 272.5 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.1 F 0.0  
5:00 231.8 F 0.0  
10:00 227.8 F 0.0  
15:00 228.6 F 0.0  
20:00 228.4 F 0.0  
25:00 226.5 F 0.0  
30:00 228.4 F 0.0  
Drying Complete

FILLING: 1:30  
HEATING: 23:24  
STERILIZING: 4:00  
VENTING: 3:49  
DRYING: 30:00  
TOTAL CYCLE: 01:02:43  
CYCLE COMPLETE

01/03/2024 04:47 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_