
Midmark M11 - v1.0.5
Total Cycles: 3409

Name: RITTER2
Sterilizer ID: V2313228
02/23/2024 04:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.3 F 0.0
2:00 124.7 F 0.5
4:00 146.0 F 2.6
6:00 170.8 F 6.1
8:00 193.7 F 10.3
10:00 224.9 F 6.8
12:00 231.0 F 8.5
14:00 241.2 F 13.1
16:00 237.4 F 10.2
18:00 236.0 F 8.9
20:00 248.7 F 15.0
22:00 260.2 F 21.4
24:00 270.7 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 29.5
1:00 273.0 F 30.3
2:00 271.7 F 29.5
3:00 272.6 F 30.3
4:00 272.1 F 29.8

Min 271.2 F 29.4
Max 273.2 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.7 F 0.0
5:00 226.2 F 0.0
10:00 221.4 F 0.0
15:00 225.7 F 0.0
20:00 220.0 F 0.0
25:00 203.2 F 0.0
30:00 191.3 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 25:26
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:04:05
CYCLE COMPLETE

02/23/2024 05:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____