
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 000528
14:24 04/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:08
PRESSURIZING 3:15
STERILIZING 4:28
136.0 C 228kPa 4:28
136.1 C 230kPa 4:58
136.3 C 232kPa 5:28
136.2 C 228kPa 5:58
136.4 C 232kPa 6:28
136.4 C 232kPa 6:58
136.3 C 232kPa 7:28
136.4 C 230kPa 7:58
136.4 C 230kPa 7:58
VENTING 7:58
AIR DRYING 8:53
CYCLE COMPLETE 11:58

Digital Signature #
E924976F03BAFE3E
