
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.7uS / 0.4ppm
CYCLE NUMBER 001684
9:02 07/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:32
PRESSURIZING 3:33
STERILIZING 4:38
136.5 C 225kPa 4:38
136.2 C 222kPa 5:08
136.3 C 220kPa 5:38
136.2 C 220kPa 6:08
136.4 C 224kPa 6:38
136.3 C 223kPa 7:08
136.3 C 221kPa 7:38
136.4 C 224kPa 8:08
136.4 C 224kPa 8:08
VENTING 8:09
AIR DRYING 9:04
CYCLE COMPLETE 9:07

Digital Signature #
D21BF79366FF4504
