
Midmark M11 - v1.0.0
Total Cycles: 4627

Name: RITTER 1
Sterilizer ID: V2178198
06/12/2024 08:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.7 F 0.0
2:00 74.6 F 0.0
4:00 86.8 F 0.1
6:00 120.3 F 0.2
8:00 148.2 F 0.9
10:00 170.9 F 3.4
12:00 185.1 F 6.1
14:00 198.4 F 9.0
16:00 209.1 F 12.1
18:00 215.3 F 4.3
20:00 227.4 F 7.8
22:00 236.8 F 11.6
24:00 233.5 F 7.8
26:00 244.3 F 13.0
28:00 242.0 F 11.2
30:00 253.2 F 16.8
32:00 263.1 F 22.6
34:00 272.6 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.7 F 28.9
2:00 271.9 F 28.1
3:00 272.1 F 28.7
4:00 272.5 F 28.7

Min 271.5 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.4 F 0.0
5:00 201.9 F 0.0
10:00 203.2 F 0.0
15:00 191.1 F 0.0
20:00 208.9 F 0.0
25:00 211.9 F 0.0
30:00 211.8 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 34:27
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:12:59
CYCLE COMPLETE

06/12/2024 09:27 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____