
Midmark M11 - v1.0.5
Total Cycles: 1040

Name: RITTER2
Sterilizer ID: V2313228
04/27/2022 08:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.5 F 0.0
2:00 177.3 F 2.2
4:00 171.9 F 5.8
6:00 201.0 F 10.9
8:00 217.5 F 5.7
10:00 233.1 F 11.3
12:00 234.2 F 9.0
14:00 239.9 F 10.0
16:00 247.1 F 14.4
18:00 259.4 F 21.2
20:00 270.4 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 272.7 F 29.8
2:00 271.6 F 29.1
3:00 272.8 F 30.0
4:00 272.1 F 29.3

Min 271.5 F 29.0
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.0 F 0.0
5:00 204.2 F 0.0
10:00 216.9 F 0.0
15:00 222.7 F 0.0
20:00 226.8 F 0.0
25:00 230.3 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 20:52
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:19
CYCLE COMPLETE

04/27/2022 09:35 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____