
Midmark M11 - v1.0.5
Total Cycles: 1840

Name: RITTER2
Sterilizer ID: V2313228
12/14/2022 01:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.9 F 0.0
2:00 169.0 F 3.8
4:00 180.6 F 8.1
6:00 204.0 F 13.5
8:00 221.4 F 7.9
10:00 234.0 F 13.1
12:00 238.5 F 11.4
14:00 238.5 F 10.3
16:00 252.7 F 17.3
18:00 265.0 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.2 F 29.3
2:00 271.8 F 29.7
3:00 271.9 F 29.2
4:00 271.8 F 29.7

Min 271.3 F 28.9
Max 272.5 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.7 F 0.0
5:00 226.9 F 0.0
10:00 224.9 F 0.0
15:00 226.4 F 0.0
20:00 223.8 F 0.0
25:00 224.4 F 0.0
30:00 222.0 F 0.0

Drying Complete

FILLING: 1:16
HEATING: 19:46
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:11
CYCLE COMPLETE

12/14/2022 02:35 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____