
Midmark M11 - v1.0.5
Total Cycles: 4104

Name: RITTER2
Sterilizer ID: V2313228
11/08/2024 03:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 88.1 F 0.0
2:00 98.8 F 0.0
4:00 135.7 F 0.2
6:00 188.6 F 2.7
8:00 208.9 F 7.6
10:00 218.8 F 13.5
12:00 226.6 F 8.2
14:00 239.8 F 13.0
16:00 240.3 F 11.2
18:00 238.6 F 10.0
20:00 253.3 F 17.3
22:00 265.8 F 24.9
24:00 267.2 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.7
1:00 271.8 F 28.6
2:00 272.0 F 28.5
3:00 272.3 F 28.8
4:00 272.7 F 29.0

Min 271.8 F 28.5
Max 272.9 F 29.5

VENTING CHAMBER

C105: VENT MODE
STOP PRESSED
CHAMBER TEMP 251F
CHAMBER PRESS 8.5PSI

** ITEMS NOT STERILE **

11/08/2024 04:04 PM