
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 008387
16:06 06/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:07
PRESSURIZING 2:11
STERILIZING 3:05
136.0 C 220kPa 3:05
136.4 C 220kPa 3:35
136.4 C 224kPa 4:05
136.4 C 224kPa 4:35
136.4 C 223kPa 5:05
136.4 C 225kPa 5:35
136.2 C 221kPa 6:05
136.2 C 219kPa 6:35
136.2 C 219kPa 6:35
VENTING 6:35
AIR DRYING 7:27
CYCLE COMPLETE 18:31

Digital Signature #
6AEFE21936DB63EF
